



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN PROTEIN C LRT
Product Code : 221211
Lot No. : FC01991K
Expiration date : 2025-07-13

Components	Quantity	Lot	Expiration Date
R1 : Protac	3 vials	FC0199a61K	2025-07-13
R2 : PC substrate	3 vials	FC0199a71K	2025-07-13

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	A405(100%) = 0.12 ± 0.04 Delta A405(100-50%) > 0.03 R ² ≥ 0.98 Measured values for controls in acceptance ranges	0.117 (CS) 0.08 (CS) PASS (CS) PASS
Detection Threshold	≤ 5%	< 1 %

Calibration Curve for information

CS						Acceptance Range	MV	
Conc.	150	100	50	0	R ²	NC	[84 - 104]	93
OD	0.170	0.117	0.061	0.000	0.999	AC	[30 - 41]	34

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-04-19

Quality Control Manager : S.LECOURT

P/O

