



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN Plasminogen LRT
Product Code: 221511
Lot No.: FC02171J
Expiration date: 2025-07-18

Components	Quantity	Lot	Expiration Date
R1 : Streptokinase	3 vials	FC0217b81J	2025-07-18
R2 : Substrate	3 vials	FC0217b91J	2025-07-18

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	A405(100%) = 0.45 ± 0.20 Delta A405 (100-50%) > 0.15 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.479 0.236 PASS PASS
Detection Threshold	< 5 %	< 1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2023-04-04

Quality Control Manager: **S.LECOURT**