



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL LOW CI
Product Code : 224301
Lot No. : FC02341G
Expiration date : 2025-07-22

Components	Quantity	Lot	Expiration Date
CI	12 vials	FC0234BQ1G	2025-07-22

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml	0.26 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-02-22

Quality Control Manager :

S.LECOURT