



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN™ Rivaroxaban Control Low  
**Product Code :** 225101  
**Lot No. :** FC01141E  
**Expiration date :** 2025-07-07

Components	Quantity	Lot	Expiration Date
CI	6 vials	FC0114DM1E	2025-07-07
CII	6 vials	FC0114DN1E	2025-07-07

Analytical Data	Specification	Result
Concentration	CI : 15 -35 ng/ml CII : 45 -105 ng/ml	26 ng/mL 82 ng/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2023-02-14

Quality Control Manager :

**S. LECOURT**