



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR LOW

Product Code : 226001

Lot No. : FC01761C

Expiration date : 2025-07-15

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FC0176DJ1C	2025-07-15
CAL II	4 vials	FC0176DK1C	2025-07-15
CAL III	4 vials	FC0176DL1C	2025-07-15

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -100 ng/ml	0 ng/mL 51 ng/mL 93 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-02-17

Quality Control Manager :

S.LECOURT

