



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CALIBRATOR

Product Code : 226201

Lot No. : FC01831D

Expiration date : 2025-07-12

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FC0183DO1D	2025-07-12
CAL 2	4 vials	FC0183DP1D	2025-07-12
CAL 3	4 vials	FC0183DQ1D	2025-07-12

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -400 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/mL 305 ng/mL 510 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-02-22

Quality Control Manager :

S.LECOURT

P/O 