HYPHEN BioMed

HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description :	**BIOPHEN DABIGATRAN CONTROL PLASMA LOW**
Product Code :	**225001**
Lot No. :	**FC084719**
Expiration date :	**2025-08-18**

Components	Quantity	Lot	Expiration Date
C1	6 vials	FC0847DD19	2025-08-18
C2	6 vials	FC0847DE19	2025-08-18

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 45 -105 ng/ml	26 ng/mL 75 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-03-23

Quality Control Manager : **S.LECOURT**