



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : GFC Test
Product Code : CK093K
Lot No. : FC0887
Expiration date : 2025-08-30

Components	Quantity	Lot	Expiration Date
R1 : tPA with silica	3 vials	FC0887	2025-10-21
R2 : Thrombin Reagent	3 vials	FC0887	2025-08-30

Analytical Data	Specification	Result
Normal Plasma	Mean : 30 - 60 min	38 min
Controls	Lysis time within acceptance range	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-05-15

Quality Control Manager :

S.LECOURT

