



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PLASMA ARGATROBAN
CALIBRATOR
Product Code : SC030K
Lot No. : FC084918
Expiration date : 2025-08-19

| Components | Quantity | Lot | Expiration Date |
|------------|----------|------------|-----------------|
| CAL 1 | 4 vials | FC0849BZ18 | 2025-08-19 |
| CAL 2 | 4 vials | FC0849CA18 | 2025-08-19 |
| CAL 3 | 4 vials | FC0849CB18 | 2025-08-19 |
| CAL 4 | 4 vials | FC0849CC18 | 2025-08-19 |
| CAL 5 | 4 vials | FC0849CD18 | 2025-08-19 |

| Analytical Data | Specification | Result |
|-----------------|------------------------------|--------|
| CAL 1 | $\leq 0.05 \mu\text{g/mL}$ | 0.00 |
| CAL 2 | 0.35 - 0.65 $\mu\text{g/mL}$ | 0.47 |
| CAL 3 | 0.85 - 1.15 $\mu\text{g/mL}$ | 0.89 |
| CAL 4 | 1.30 - 1.65 $\mu\text{g/mL}$ | 1.39 |
| CAL 5 | 1.70 - 2.00 $\mu\text{g/mL}$ | 1.78 |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-03-14

Quality Control Manager : **S.LECOURT**