



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : GFC Control Plasmas
Product Code : SC104K
Lot No. : FC0725
Expiration date : 2025-08-16

Components	Quantity	Lot	Expiration Date
C1	2 vials	FC0725	2025-08-16
C2	2 vials	FC0725	2025-08-16
C3	2 vials	FC0725	2025-08-16

Analytical Data	Specification	Result
C1	10 -20 min	12 min
C2	35 - 55 min	35 min
C3	65 -110 min	71 min

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-03-16

Quality Control Manager : **S.LECOURT**

