



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CONTROL LOW
Product Code : 225401
Lot No. : FC10851A
Expiration date : 2025-09-27

Components	Quantity	Lot	Expiration Date
CI	6 vials	FC1085ai1A	2025-09-27
CII	6 vials	FC1085aj1A	2025-09-27

Analytical Data	Specification	Result
Concentration	CI : 15 -35 ng/ml CII : 45 -105 ng/ml	28 ng/mlL 81 ng/mlL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-05-15

Quality Control Manager :

S.LECOURT