



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CALIBRATOR
Product Code : 226501
Lot No. : FC10431C
Expiration date : 2025-09-23

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FC1043aa1C	2025-09-23
CAL 2	4 vials	FC1043ab1C	2025-09-23
CAL 3	4 vials	FC1043ac1C	2025-09-23

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 40 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 241 480
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-05-15

Quality Control Manager :

S.LECOURT

