

Art no: 10-0027/10-0028

CERTIFICATE OF ANALYSIS

Lot no: 301105



Intertek


Product name:	Protein L ligand leakage ELISA kit		
Production date:	2024-04		
Date of release:	2024-06-10		
Expiry date:	2025-10		
Store:	+4°C. Protein L reference at -20°C or -80°C.		
Contents:	ELISA kit for detection and quantification of ligand leakage in eluates from Protein L affinity chromatography medium containing immunoglobulin (Ig) or Ig-fragments.		
Kit contents	Lot	Expiry date	
1 x Anti-Protein L coated microtiter plate	301507	2025-10	
130 µl HRP-conjugated anti-Protein L pAb	301904	2025-10	
1 x 200 µl Protein L reference (600 ng/ml) (10-0027)	285004	2025-10	
5 x 200 µl Protein L reference (600 ng/ml) (10-0028)	285004	2025-10	
13 ml EC-Blue Enhanced™	301107	2025-10	
7 ml Stop solution, 0.5 M H ₂ SO ₄	301106	2025-10	
6 x 5 g PBS Tween tablets	292707-03	2026-07	

Quality Control

Analysis	Specification	Result
%CV	< 12%	Passed
Absorbance 30 ng/ml	> 2.1	Passed

	OD 450 nm
Calibrator 0.47 ng/ml	0.126
Calibrator 0.94 ng/ml	0.194
Calibrator 1.88 ng/ml	0.369
Calibrator 3.75 ng/ml	0.717
Calibrator 7.5 ng/ml	1.403
Calibrator 15 ng/ml	2.232
Calibrator 30 ng/ml	2.529

	Concentration
Control sample A	0.157
Control sample B	0.507
Control sample C	1.545



The above material has met all quality specifications and has been reviewed by a quality representative

Cynthia Hassinen

2024-06-10

Quality Assurance, Cynthia Hassinen

Date