



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

**Item Description:** LIAPHEN Fibrinogen

**Item Code:** 120102

**Lot No.:** FC13631J

**Expiration date:** 2025-11-29

Components	Quantity	Lot	Expiration Date
R1: Latex reagent	4 vials	FC1363JA1J	2025-11-29

Analytical Data	Specification	Result
Calibration curve: reactivity, slope, linearity and accuracy	Delta A620 (0-20µg/ml) ≥ 0.40 Controls measured within acceptance ranges R <sup>2</sup> ≥ 0.98	0.57 PASS 0.99
Limit of Quantification	≤ 1 µg/ml (in the test dilution (1:200))	< 1
Hook effect	> 12 g/L (in the test dilution (1:200))	> 12

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2023-07-05

Quality Control Manager: **S. LECOURT**