



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS**Product Description : ZYMUTEST TOTAL TF****Product Code : RK042A****Lot No. : FC1341****Expiration date : 2025-11-07**

Components	Quantity	Lot	Expiration Date
Anti-(h)-TF pre-coated plate	12x8 wells	FC1288	2025-11-08
TF Standard	3 vials	FC1301	2025-11-09
Biotinylated Antibody	1 vial	FC1299	2025-11-15
Control I	1 vial	FC1302	2025-11-09
Control II	1 vial	FC1303	2025-11-09
HRP streptavidine immunoconjugate	3 vials	FC1364	2025-11-17
TF-Sample diluent	2x40 ml	FC1290	2025-11-07
Wash solution	1x50 ml	FC0035	2026-06-15
TF - Conjugate diluent	1x25 ml	FC1289	2025-11-07
TMB HS	1x25 ml	220103D06	2026-01-31
Sulfuric Acid 0,45M	1x6 ml	FB1450	2026-02-23

Analytical Data	Specification	Result
Reactivity	A450 for 500 pg/ml ≥ 1.60	1.30 (*)
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.150	0.145
Detection threshold	≤ 10 pg/ml	$= 10$ pg/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

() accepted under waiver NCI n° 2180*

Date of Approval : 2023-06-13

Quality Control Manager : **S. LECOURT**