



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

CERTIFICATE OF ANALYSIS	
Product Description :	LA CONTROL PLASMA HIGH
Product Code :	SC083K
Lot No. :	FC21251L
Expiration date :	2025-11-29

Components	Quantity	Lot	Expiration Date
C2	6 vials	FC2125g81L	2025-11-29

Analytical Data	Specification	Result
C2	Normalized Ratio : $\geq 2.00$	STAR : 2.31 CS : 2.52 CN : 2.54

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval** Quality Control Manager :

10 OCT. 2023

S. LECOURT