


HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

CERTIFICATE OF ANALYSIS	
<b>Product Description:</b>	<b>Factor VIII:C Deficient Plasma</b>
<b>Product Code:</b>	<b>DP040A/K</b>
<b>Lot No.:</b>	<b>FC15861K</b>
<b>Expiration date:</b>	<b>2025-12-20</b>

Components	Quantity	Lot	Expiration Date
Factor VIII:C Deficient Plasma	1 vial (DP040A) 6 vials (DP040K)	FC1586JV1K	2025-12-20

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) ≥40 sec CT (100%) <75 sec	84sec 59sec
Concentration FVIII	< 1%	<1%
Concentration FV	> 50 %	71%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: **Quality Control Approval** Quality Control Manager: 

**20 JUL. 2023**  
**S. LECOURT**