



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN AT

Item Code : 120002

Lot No. : FC1625FG

Expiration date : 2026-01-09

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	2 vials	FC1625IWFG	2026-01-09
R2 : Reaction Buffer (*)	2 vials	FC1625IXFG	2026-01-09

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	A620 (0-150%) ≥ 0.40 $R^2 \geq 0.98$ Controls measured within acceptance ranges	0.66 0.99 PASS
Detection threshold	$\leq 10\%$	< 5%
Hook effect	>250%	PASS

(*) Accepted under waiver (NCI N°2207)

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-09-14

Quality Control Manager : **S.LECOURT**

