



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description	BIOPHEN PLASMA CALIBRATOR
Product Code :	222101
Lot No. :	FC17031B
Expiration date :	2026-01-10

Components	Quantity	Lot	Expiration Date
Plasma Calibrator	12 vials	FC1703DY1B	2026-01-10

Analytical Data	Specification	Result
Fibrinogen (Activity)	> 2 g/L	2.61
Fibrinogen (Antigen)	> 2 g/L	2.74
Alpha 2 Anti plasmin	> 85 %	97%
Antithrombin (Activity)	> 85 %	91%
Antithrombin (Antigen)	> 85 %	89%
Factor II	> 80 %	97%
Factor V	> 80 %	104%
Factor VII	> 80 %	110%
Factor X	> 80 %	92%
Factor VII+ X	> 85 %	103%
Factor VIII	> 60 %	98%
Factor IX	> 60 %	100%
Factor XI	> 70 %	90%
Factor XII	> 70 %	110%
Factor XIII	> 70 %	85%
Free Protein S (Antigen)	70 - 140 %	131%
Plasminogen	> 80 %	105%
Protein C Chromogenic	> 85 %	101%
Protein C Clotting	> 70 %	106%
Protein C (Antigen)	> 75 %	88%
Protein S (Activity)	> 50 %	106%
von Willebrand Factor (Antigen)	> 60 %	97%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-10-19

Quality Control Manager : S. Lecourt

