



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR

Product Code : 222701

Lot No. : FC17441D

Expiration date : 2026-01-20

| Components | Quantity | Lot | Expiration Date |
|------------|----------|------------|-----------------|
| CAL 1 | 4 vials | FC1744BA1D | 2026-01-20 |
| CAL 2 | 4 vials | FC1744BB1D | 2026-01-20 |
| CAL 3 | 4 vials | FC1744BC1D | 2026-01-20 |

| Analytical Data | Specification | Result |
|--|---|--------------------------------|
| Concentration | CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -500 ng/ml | 0ng/mL 265ng/mL 472ng/mL |
| Linearity and accuracy of calibration with values assigned | $R^2 \geq 0.98$ Measured values for controls in acceptance ranges | PASS PASS |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-10-20

Quality Control Manager : S.LECOURT

