



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description BIOPHEN ABNORMAL CONTROL PLASMA

Product Code : 223301

Lot No. : FC17081A

Expiration date : 2026-01-23

Components	Quantity	Lot	Expiration Date
Abnormal Control	12 vials	FC1708EA1A	2026-01-23

Analytical Data	Specification	Result
INR	> 1,3	>1.3
aPTT LR Ratio	> 1.50	>1.50
Fibrinogen (Activity)	0.8-1.2 g/L	1.01g/L
Fibrinogen (Antigen)	0.7-1.3 g/L	1.12g/L
LA ratio	≤ 1.2	<1.2
Alpha 2 Anti plasmin	20 - 50 %	35%
Antithrombin (Activity)	20 - 50 %	35%
Antithrombin (Antigen)	20 - 50 %	34%
Factor II	20 - 50 %	36%
Factor V	20 - 50 %	30%
Factor VII	20 - 50 %	34%
Factor X	20 - 50 %	32%
Factor VII+ X	20 - 50 %	34%
Factor VIII	20 - 50 %	30%
Factor IX	20 - 50 %	32%
Factor XI	20 - 50 %	30%
Factor XII	20 - 50 %	36%
Factor XIII	20 - 50 %	46%
Free Protein S (Antigen)	20 - 50 %	33%
Plasminogen	20 - 50 %	37%
Protein C Chromogenic	20 - 50 %	35%
Protein C (Antigen)	20 - 50 %	31%
Protein C Clotting	20 - 50 %	41%
Protein S (Activity)	20 - 50 %	35%
von Willebrand Factor (Antigen)	20 - 50 %	37%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-10-24

Quality Control Manager : S.LECOURT