



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** Calcium Chloride 0.025M (M/40)  
**Product Code:** AR001B/L  
**Lot No.:** FC1720FL  
**Expiration date:** 2026-01-04

Components	Quantity	Lot	Expiration Date
Calcium Chloride 0.025M (M/40)	3 vials (AR001B) 12 vials (AR001L)	FC1720j2FL	2026-01-04

Analytical Data	Specification	Result
aPTT on citrated human plasma (CEPHEN)	ratio $\leq$ 1.20	1.01
FVIII:C clotting assay	Delta CT (100%) with control sample $\leq$ 5%	< 1%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2023-08-21

Quality Control Manager: S. LECOURT