



 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	Prothrombin Deficient Plasma
Product Code :	DP010A/K
Lot No. :	FC17421E
Expiration date :	2026-01-19

Components	Quantity	Lot	Expiration Date
Prothrombin Deficient Plasma	1 vial (DP010A) 6 vials (DP010K)	FC1742JO1E	2026-01-19

Analytical Data	Specification	Result
Calibration curve	CT (0%) >100 sec CT (100%) <30 sec Thromboplastin: STAGO	>250sec 27sec
Concentration FII	< 1%	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

21 AOUT 2023

S. LECOURT