



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST (h) PAI-1 Antigen

Product Code : RK012A

Lot No. : FC1069

Expiration date : 2026-01-31

Components	Quantity	Lot	Expiration Date
Anti-(h)-PAI-1 pre-coated plate	12x8 wells	FC1214	2026-10-13
(h) PAI-1 standard	3 vials	FC1246	2026-10-22
(h) PAI-1 Control I	1 vial	FC1185	2026-10-08
(h) PAI-1 Control II	1 vial	FC1186	2026-10-08
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	FC1287	2026-11-02
F-Sample diluent	2x50 ml	FC1248	2026-10-20
Wash solution	1x50 ml	FC0034	2026-06-15
Conjugate diluent	1x25 ml	FC0128	2026-06-30
TMB	1x25 ml	220106D03	2026-01-31
Sulfuric Acid 0,45M	1x6 ml	FC0223	2026-07-15

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml \geq 1.40	1.59
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.004
Detection threshold	\leq 0.5 ng/ml	< 0.1 ng/mL

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20	Acceptance Range	
						CI	MV
OD	1.856	1.245	0.714	0.332	0.172	[34.0 - 46.1]	40.0
						[15.3 - 22.9]	18.5

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-06-02

Quality Control Manager :

S.LECOURT

