



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

**CERTIFICATE OF ANALYSIS**
**Product Description : ZYMUTEST HIA IgG,A,M**
**Product Code : RK040D**
**Lot No. : FC2291**
**Expiration date : 2026-02-23**

Components	Quantity	Lot	Expiration Date
Heparin pre-coated plate	12x8 wells	FC2321	2026-05-01
HIA Ig G,A,M Positive Control	3 vials	FC2390	2026-05-18
Negative Control	3 vials	FC2275	2027-04-23
Anti (h) IgG - Anti IgA - Anti IgM HRP immunoconjugate	3 vials	FC2489	2027-05-27
HIA Sample diluent	2x50 ml	FC2273	2026-04-24
Wash solution	1x50 ml	FD0073	2027-06-22
Cell Lysate	3 vials	FC2364	2026-05-08
Conjugate diluent	1x25 ml	FC1583	2026-12-16
TMB	1x25 ml	221201D01	2026-12-31
Sulfuric Acid 0,45M	1x6 ml	FB1449	2026-02-23

Analytical Data	Specification	Result
Reactivity	A450 for Positive Control $\geq 1.00$ A450 for Negative Control $\leq 0.30$	1.918 0.138
Pathological Plasmas	A450 $\geq 0.500$	PASS
Blank	$< 0.100$	0.021
Normal Plasma	M + 2SD $\leq 0.50$	0.368

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

**Quality Control Approval**

Date of Approval :

16 JAN. 2024

Quality Control Manager :

S. LECOURT

