

 HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ARGATROBAN CONTROL PLASMA  
**Product Code :** SC035K  
**Lot No. :** FC190819  
**Expiration date :** 2026-02-14

Components	Quantity	Lot	Expiration Date
C1	6 vials	FC1908CE19	2026-02-14
C2	6 vials	FC1908CF19	2026-02-14

Analytical Data	Specification	Result
C1	0.55 - 0.80 µg/mL	0.69 µg/ml
C2	1.10 - 1.40 µg/mL	1.29 µg/ml

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

Quality Control Manager :

Quality Control Approval

18 SEP. 2023  
S. LECOURT

