



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN FACTOR X

**Product Code :** 221705

**Lot No. :** FC21471B

**Expiration date :** 2026-03-27

Components	Quantity	Lot	Expiration Date
R1 : Substrate	4 vials	FC21471B	2026-03-31
R2 : RVV	4 vials	FC21471B	2026-04-05
R3: Buffer	4 vials	FC21471B	2026-03-27

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	$A_{405} (1:5) \geq 0.75$ $\Delta A_{405} (1:5 ; 1:10) \geq 0.35$ $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.03 0.50 1.00 PASS
Detection Threshold	$\leq 5\%$	$< 1\%$

### Calibration Curve for information

STAR

Conc.	0	50	100	200
OD	0	0.27	0.53	1.03

	Acceptance Range	MV
NC	[ 70 - 90 ]	82
AC	[ 25 - 34 ]	32

MV = Measured value

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2023-11-07

Quality Control Manager : **S. LECOURT**

