



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN HEPARIN CALIBRATOR
Product Code : 222001
Lot No. : FC2150GC
Expiration date : 2026-03-28

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FC2150BJGC	2026-03-28
CAL 2	4 vials	FC2150BKGC	2026-03-28
CAL 3	4 vials	FC2150BLGC	2026-03-28
CAL 4	4 vials	FC2150BMGC	2026-03-28
CAL 5	4 vials	FC2150BNGC	2026-03-28

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 0.05 IU/ml CAL 2 : 0.30 -0.50 IU/ml CAL 3 : 0.70 -0.90 IU/ml CAL 4 : 1.10 -1.30 IU/ml CAL 5 : 1.45 -1.75 IU/ml	0.00 IU/mL 0.41 IU/mL 0.79 IU/mL 1.17 IU/mL 1.48 IU/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	Pass Pass

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-10-23

Quality Control Manager : S.LECOURT

