



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL LOW CI
Product Code : 224301
Lot No. : FC2112G6
Expiration date : 2026-03-21

Components	Quantity	Lot	Expiration Date
CI	12 vials	FC2112BQG6	2026-03-21

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml	0.30IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-10-23

Quality Control Manager : S.LECOURT