



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN EDOXABAN CALIBRATOR LOW  
**Product Code :** 226401  
**Lot No. :** FC09951A  
**Expiration date :** 2026-03-03

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FC0995af1A	2026-03-03
CAL II	4 vials	FC0995ag1A	2026-03-03
CAL III	4 vials	FC0995ah1A	2026-03-03

Analytical Data	Specification	Result
Concentration	CAL I : < 15 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 mg/ml 48 mg/ml 105 mg/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2023-12-11

Quality Control Manager : S. LECOURT

