



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL PLASMA
Product Code : 223101
Lot No. : FC2472 / FC2548
Expiration date : 2026-04-18

Components	Quantity	Lot	Expiration Date
C1	6 vials	FC2253BXGV	2026-04-18
C2	6 vials	FC2249BYGU	2026-04-18

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml C2 : 0.40 -0.60 IU/ml	0.20 IU/mL 0.54 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-12-19

Quality Control Manager :

S.LECOURT