



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : HEMOCLOT™ LA-C
Product Code : CK091K
Lot No. : FC22701G
Expiration date : 2026-04-24

Components	Quantity	Lot	Expiration Date
HEMOCLOT LA-C	6 vials	FC2270g61G	2026-04-24

Analytical Data	Specification	Result
Normal Plasma	Mean CT < 45 sec NR < 1.20	CN : 33.5 sec STAR : 35.9 sec CS : 33.2 sec PASS
Pathological Plasma	NR ≥ 1.20	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-12-05

Quality Control Manager : S. LECOURT