



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ V-L Control Plasma  
Product Code : 223405  
Lot No. : FC24931B  
Expiration date : 2026-05-30

Components	Quantity	Lot	Expiration Date
V-L Control Plasma	12 vials	FC2493FX1B	2026-05-30

Analytical Data	Specification	Result
% FVL	25-75 %	36%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2024-01-05

Quality Control Manager : S.LECOURT