



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : CEPHEN 5
Product Code : CK515K/L
Lot No. : FD0968KA
Expiration date : 2026-05-18

Components	Quantity	Lot	Expiration Date
CEPHEN 5	8 vials (CK515K) 12 vials (CK515L)	FD0968KBKA	2026-05-18

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	1.02 on STAR 1.01 on CS 1.01 on CN

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

11 JUIN 2024

S. LECOURT

Quality Control Manager :

