



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description: LIAPHEN™ Protein C

Item Code: 120003

Lot No.: FD212619 / FD257312

Expiration date: 2026-06-30

Components	Quantity	Lot	Expiration Date
R1: Reaction Buffer	4 vials	FD2126en19 / FD257312	2026-07-21
R2: Latex reagent	4 vials	FD2126eo19 / FD257312	2026-06-30

Analytical Data	Specification	Result
Calibration curve : reactivity, slope and accuracy	OD/min (3/2) \geq 0.30 Delta OD/min (1/1-1/2) \geq 0.1 Controls measured within acceptance ranges	0.45 0.15 PASS
Hook effect	OD/min (300 et 600 %) \geq OD/min (point 1/1 of calibration curve)	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval:

Quality Control Approval Quality Control Manager:

17 MARS 2025

S. LECOURT