



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	BIOPHEN™ BIVALIRUDIN CALIBRATOR
Product Code :	226701
Lot No. :	FC255216 / FC255114
Expiration date :	2026-06-27

Components	Quantity	Lot	Expiration Date
CAL 1	2 vials	FC2552cz16 / FC255114	2026-06-27
CAL 2	2 vials	FC2552da16 / FC255114	2026-06-27
CAL 3	2 vials	FC2552db16 / FC255114	2026-06-27
CAL 4	2 vials	FC2552dc16 / FC255114	2026-06-27
CAL 5	2 vials	FC2552dd16 / FC255114	2026-06-27

Analytical Data	Specification	Result
Concentration	CAL 1 : < 0.25 µg/mL CAL 2 : 0.8 -1.2 µg/mL CAL 3 : 1.6 -2.4 µg/mL CAL 4 : 3.0 - 4.0 µg/mL CAL 5 : 4.0 - 5.0 µg/mL	0 µg/mL 1.07 µg/mL 2.19 µg/mL 3.90 µg/mL 5.00 µg/mL
Accuracy of calibration with values assigned	Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval : 13 FEB. 2024 Quality Control Manager :

S. LECOURT

