



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: HEMOCLOT FACTEUR VIIa
Product Code: CK092K
Lot No.: FC16021C
Expiration date: 2026-06-17

Components	Quantity	Lot	Expiration Date
R1: Factor VII Deficient Plasma	3 vials	FC1602ak1C	2026-06-27
R2 : rTTF and PLPs	3 vials	FC1602al1C	2026-07-04
R3 : Hepes BSA Buffer	3 vials	FC1602am1C	2026-06-17

Analytical Data	Specification	Result
Calibration curve on CS	Delta CT (500-12.5 mIU/mL) ≥ 60 sec	91,6 sec
	CT (250 mIU/mL) = 40 ± 7 sec	42,1 sec
	$R^2 \geq 0.98$	PASS
	Normal Plasma : 15-140 mIU/mL	Normal Plasma : 58 mIU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2023-08-16

Quality Control Manager:

S. LECOURT