



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN RIVAROXABAN CONTROL PLASMA  
**Product Code :** 224501  
**Lot No. :** FD02051F  
**Expiration date :** 2026-07-20

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD0205BD1F	2026-07-20
C2	6 vials	FD0205BE1F	2026-07-20

Analytical Data	Specification	Result
Concentration	C1 : 50 -150 ng/ml C2 : 250 -350 ng/ml	96 ng/mL 307 ng/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

### Quality Control Approval

Date of Approval :

**07 MARS 2024**

**S. LECOURT**

Quality Control Manager :