



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN APIXABAN CALIBRATOR LOW  
**Product Code :** 226101  
**Lot No. :** FD02821D  
**Expiration date :** 2026-07-27

| Components | Quantity | Lot        | Expiration Date |
|------------|----------|------------|-----------------|
| CAL I      | 4 vials  | FD0282DT1D | 2026-07-27      |
| CAL II     | 4 vials  | FD0282DU1D | 2026-07-27      |
| CAL III    | 4 vials  | FD0282DV1D | 2026-07-27      |

| Analytical Data  | Specification  | Result                           |
|--|--|----------------------------------|
| Concentration  | CAL I : < 20 ng/ml<br>CAL II : 40 -60 ng/ml<br>CAL III : 80 -120 ng/ml | 0 ng/ml<br>48 ng/ml<br>100 ng/ml |
| Linearity and accuracy of calibration with values assigned | $R^2 \geq 0.98$<br>Measured values for controls in acceptance ranges   | PASS                             |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

Quality Control Manager :

Quality Control Approval

29 FEV. 2024  
S. LECOURT