

Product name: Protein L ligand leakage ELISA kit

Production date: 2024-04
Date of release: 2025-08-28 (Retested 2025-08-22)
Expiry date: 2026-08
Store: +4°C. Protein L reference at -20°C or -80°C.
Contents: ELISA kit for detection and quantification of ligand leakage in eluates from Protein L affinity chromatography medium containing immunoglobulin (Ig) or Ig-fragments.


Kit contents	Lot	Expiry date
1 x Anti-Protein L coated microtiter plate	301507	2026-08
130 µl HRP-conjugated anti-Protein L pAb	301904	2026-08
1 x 200 µl Protein L reference (600 ng/ml) (10-0027)	285004	2026-08
5 x 200 µl Protein L reference (600 ng/ml) (10-0028)	285004	2026-08
13 ml EC-Blue Enhanced™	301413-01	2026-08
7 ml Stop solution, 0.5 M H ₂ SO ₄	301106	2026-08
6 x 5 g PBS Tween tablets	300901-04	2027-02

Quality Control

Analysis	Specification	Result
%CV	< 12%	Passed
Absorbance 30 ng/ml	> 2.1	Passed

	OD 450 nm
Calibrator 0.47 ng/ml	0.145
Calibrator 0.94 ng/ml	0.243
Calibrator 1.88 ng/ml	0.460
Calibrator 3.75 ng/ml	0.856
Calibrator 7.5 ng/ml	1.574
Calibrator 15 ng/ml	2.281
Calibrator 30 ng/ml	2.596

		Concentration
Control sample A	0.203	0.76 ng/ml
Control sample B	0.623	2.60 ng/ml
Control sample C	1.509	7.20 ng/ml



The above material has met all quality specifications and has been reviewed by a quality representative

Sebastian Määtä
Quality Assurance, Sebastian Määtä

2025-08-28

Date