



HYPHEN BioMed  
 155 rue d'Eragny  
 95000 Neuville-sur-Oise  
 France

**CERTIFICATE OF ANALYSIS**

**Product Description :** BIOPHEN™ V-L Pasma Calibrator  
**Product Code :** 222401  
**Lot No. :** FD049112  
**Expiration date :** 2026-08-22

Components	Quantity	Lot	Expiration Date
CAL 1	3 vials	FD0491FU12	2026-08-22
CAL 2	3 vials	FD0491FV12	2026-08-22
CAL 3	3 vials	FD0491FW12	2026-08-22

Analytical Data	Specification	Result
Concentration	CAL 1 : 5-15% CAL 2 : 20-30% CAL 3 : 40-60%	14% 30% 58%
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**  
**25 MARS 2024**  
**S. LECOURT**  
 Quality Control Manager :

