



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ARACHIDONIC ACID
Product Code : AG003K
Lot No. : FE05393L / FE0571
Expiration date : 2026-08-26

Components	Quantity	Lot	Expiration Date
ARACHIDONIC ACID	3 vials	FE05393L / FE0571	2026-08-26

Analytical Data	Specification	Result
Reactivity (% AgregMax) 1.0 mM on CS or CN series	$\geq 50\%$	90 %
Reactivity (% AgregMax) 1.0 mM on Agregometer	$\geq 50\%$	76%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval
Date of Approval :

18 MARS 2025

S. LECOURT

Quality Control Manager :

