



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL PLASMA
Product Code : 223101
Lot No. : FD0933
Expiration date : 2026-09-22

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD0707BXJS	2026-09-22
C2	6 vials	FD0708BYJT	2026-09-22

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml C2 : 0.40 -0.60 IU/ml	0.19 IU/mL 0.54 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**
29 AVR. 2024
S. LECOURT
Quality Control Manager :