



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN DABIGATRAN CONTROL PLASMA LOW
Product Code : 225001
Lot No. : FD059222
Expiration date : 2026-09-08

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD0592DD22	2026-09-08
C2	6 vials	FD0592DE22	2026-09-08

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 40 -80 ng/ml	26 ng/mL 68 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

09 AVR. 2024

S. LECOURT