



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Calcium Chloride 0.025M
Product Code : AR001B/L
Lot No. : FD0855K3
Expiration date : 2026-10-09

Components	Quantity	Lot	Expiration Date
Calcium Chloride 0.025M	3 vials (AR001B) 12 vials (AR001L)	FD0855j2K3	2026-10-09

Analytical Data	Specification	Result
aPTT on citrated human plasma (CEPHEN)	ratio ≤ 1.20	1.02
FVIII:C clotting assay	Delta CT (100%) with control sample $\leq 5\%$	< 1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

14 MAI 2024

S. LECOURT

Quality Control Manager :