



 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description: LIAPHEN™ Protein C

Item Code: 120003

Lot No.: FE29681B / FE2988

Expiration date: 2026-12-29

Components	Quantity	Lot	Expiration Date
R1: Reaction Buffer	4 vials	FE2968en1B / FE2988	2026-12-29
R2: Latex reagent	4 vials	FE2968eo1B / FE2988	2027-01-27

Analytical Data	Specification	Result
Calibration curve : reactivity, slope and accuracy	OD/min (3/2) \geq 0.30 Delta OD/min (1/1-1/2) \geq 0.1 Controls measured within acceptance ranges	0.44 0.17 PASS
Hook effect	OD/min (300 et 600 %) \geq OD/min (point 1/1 of calibration curve)	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: **Quality Control Approval**

Quality Control Manager:

01 SEP. 2025

S. LECOURT