



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN AT (Anti Ila)
Product Code : 221122
Lot No. : FD131822
Expiration date : 2026-12-28

Components	Quantity	Lot	Expiration Date
R1 : Bovin Thrombin	2 vials	FD1318IT22	2027-01-10
R2 : Substrate	2 vials	FD1318IU22	2026-12-28
R3: Tris heparin Buffer	2 vials	FD1318IV22	2027-01-08

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	Delta A405 (0-100%) \geq 0.80 $R^2 \geq$ 0.98 Measured values for controls in acceptance ranges	1.512 PASS PASS
Detection Threshold	\leq 10%	< 5 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

11 SEP. 2024

S. LECOURT