



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN PLASMA CALIBRATOR

Product Code : 222101

Lot No. : FD1198KU

Expiration date : 2026-12-12

Components	Quantity	Lot	Expiration Date
Plasma Calibrator	12 vials	FD1198DYKU	2026-12-12

Analytical Data	Specification	Result
Fibrinogen (Activity)	> 2 g/L	PASS
Fibrinogen (Antigen)	> 2 g/L	PASS
Alpha 2 Anti plasmin	> 85 %	PASS
Antithrombin (Activity)	> 85 %	PASS
Antithrombin (Antigen)	> 85 %	PASS
Factor II	> 80 %	PASS
Factor V	> 80 %	PASS
Factor VII	> 80 %	PASS
Factor X	> 80 %	PASS
Factor VII+ X	> 85 %	PASS
Factor VIII	> 60 %	PASS
Factor IX	> 60 %	PASS
Factor XI	> 70 %	PASS
Factor XII	> 70 %	PASS
Factor XIII	> 70 %	PASS
Free Protein S (Antigen)	70 - 140 %	PASS
Plasminogen	> 80 %	PASS
Protein C Chromogenic	> 85 %	PASS
Protein C Clotting	> 70 %	PASS
Protein C (Antigen)	> 75 %	PASS
Protein S (Activity)	> 50 %	PASS
von Willebrand Factor (Antigen)	> 60 %	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

Quality Control Manager :

17 OCT. 2024

S. LECOURT