



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

<b>Product Description :</b>	<b>Factor IX Deficient Plasma</b>
<b>Product Code :</b>	<b>DP050A/K</b>
<b>Lot No. :</b>	<b>FD1357LG</b>
<b>Expiration date :</b>	<b>2026-12-28</b>

Components	Quantity	Lot	Expiration Date
Factor IX Deficient Plasma	1 vial (DP050A) 6 vials (DP050K)	FD1357JWLG	2026-12-28

Analytical Data	Specification	Result
Calibration curve	CT (0%) $\geq$ 40 SEC CT (100%) <80 SEC	166sec 54sec
Concentration FIX	< 1%	<1%
Concentration FV	> 50 %	78%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**      Quality Control Manager :

**22 JUL. 2024**

**S. LECOURT**

