



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST (h) PAI-1 Antigen

Product Code : RK012A

Lot No. : FD0330

Expiration date : 2026-12-31

Components	Quantity	Lot	Expiration Date
Anti-(h)-PAI-1 pre-coated plate	12x8 wells	FD0612	2027-09-06
(h) PAI-1 standard	3 vials	FD0566	2027-09-02
(h) PAI-1 Control I	1 vial	FD0563	2027-09-02
(h) PAI-1 Control II	1 vial	FD0561	2027-09-02
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	FD0611	2027-09-14
F-Sample diluent	2x50 ml	FD0480	2027-08-16
Wash solution	1x50 ml	FD0073	2027-06-22
Conjugate diluent	1x25 ml	FD0390	2027-08-02
TMB	1x25 ml	221201D01	2026-12-31
Sulfuric Acid 0,45M	1x6 ml	FD0463	2027-08-09

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml \geq 1.40	1.46
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.003
Detection threshold	\leq 0.5 ng/ml	<0.5 ng/ml

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20	Acceptance Range	MV
OD	1.61	1.08	0.63	0.27	0.15	[34.2 - 46.2]	43.2
						[9.0 - 13.6]	10.9

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : _____ Quality Control Manager : _____

Quality Control Approval

02 MAI 2024
S. LECOURT

